



Consumer and  
Corporate Affairs Canada

Consommation  
et Corporations Canada

Legal Metrology

Méetrologie légale

APPROVAL No. - N° D'APPROBATION  
S.WA-3310

OCT 23 1986

**NOTICE OF APPROVAL**

**AVIS D'APPROBATION**

Issued by statutory authority of  
the Minister of Consumer and Corporate  
Affairs under application by:

Accordée en vertu du pouvoir  
statutaire du Ministre de Consommation  
et Corporations à la demande de:

Pitney Bowes of Canada Limited  
150 Ferrand Drive  
Don Mills, Ontario  
M3C 3B5

for the following devices:

pour les appareils suivants:

DEVICE TYPE /  
TYPE D'APPAREIL:

MANUFACTURER /  
FABRICANT:

9 Electronic Postal Scale / Balance  
postale électronique.

Tokyo Electric Co. Limited  
1-14-10 Uchikanda, Chiyoda Ku  
Tokyo, Japan/Japon

MODEL DESIGNATIONS /  
DÉSICNATIONS DES MODÈLES:

RATING-CAPACITY-RANGE(S) /  
CLASSEMENT-CAPACITÉ-ÉTENDUE(S):

A500

500 g x 5 g.

NOTE: This approval applies only to  
devices, the design, composition,  
construction and performance of which  
are, in every material respect,  
identical to that described in the  
information submitted and are typified  
by the sample(s) submitted by the  
applicant for evaluation for approval  
in accordance with sections 14 and 15  
of the Weights and Measures Regula-  
tions. The following is a summary of  
salient features only.

REMARQUE: La présente approbation ne  
vise que les appareils dont la concep-  
tion, la composition, la construction  
et le rendement sont identiques, en  
tout point, à ceux qui sont décrits  
dans la documentation reçue et pour  
lesquels des échantillons représenta-  
tifs ont été fournis par le requérant  
aux fins d'évaluation, conformément aux  
articles 14 et 15 du Règlement sur les  
poids et mesures. Ce qui suit est une  
brève description de leurs principales  
caractéristiques.

## SUMMARY DESCRIPTION:

This device is an electronic postal scale designed to be used in a mail room only. The scale has 5 gram division capability which determines the correct rates by the 17 annunciators and the weight chart provided on the front of the scale.

This device is equipped with a model LC-A31 load cell which has a rated capacity of 500g.

The use of this scale is restricted to postage determination and is not permitted in direct sales to the public.

The enclosure can be sealed externally, however, the design is such that it cannot be sealed as per the requirements of SCM3/10.

## DESCRIPTION SOMMAIRE:

Il s'agit d'une balance postal électronique conçue pour être utilisée uniquement dans une salle de courrier. La balance comporte une fonction permettant de déterminer les tarifs exacts par multiples de 5 grammes grâce aux 17 voyants et à l'indicateur pondéral se trouvant à l'avant de la balance.

Cette balance est pourvue d'une cellule de pesage de modèle LC-A31 d'une capacité de 500g.

La présente balance est réservée à la détermination des tarifs postaux, et est interdite pour la vente directe au public.

Le boîtier peut être scellé de l'extérieur; toutefois, il ne peut pas être scellé conformément aux exigences de la directive ministérielle SGM3/10.

## APPROVAL:

The design, composition, construction and performance of the device type(s) identified herein having been evaluated in accordance with regulations and specifications relating thereto, established under the Weights and Measures Act, approval is hereby granted pursuant to subsection 3(1) of the said Act.

The marking, installation, use and manner of use in trade of devices are subject to inspection in accordance with regulations and specifications relating thereto, established under the Weights and Measures Act, and certification of conformity is required in addition to this approval. Inquiries regarding inspection and certification of conformity should be addressed to the local inspection office of Consumer and Corporate Affairs Canada. Requirements relating to marking are set forth in sections 18 to 26 of the Weights and Measures Regulations. Requirements relating to installation, use and manner of use are set forth in Part V and in specifications established pursuant to section 27 of the said Regulations.

## APPROBATION:

La conception, la composition, la construction et le rendement des types d'appareils identifiés ci-dessus ayant fait l'objet d'une évaluation aux fins d'approbation conformément aux Règlements et aux prescriptions établis en vertu de la Loi sur les poids et mesures, une approbation est accordée par les présentes en application du paragraphe 3(1) de ladite loi.

Le marquage, l'installation, l'utilisation et le mode d'emploi des appareils sont soumis à l'inspection conformément aux Règlements et aux prescriptions établis en vertu de la Loi sur les poids et mesures, et doivent être certifiés conformes en sus d'être approuvés par les présentes. Toute demande de renseignements sur l'inspection et la certification de conformité doit être adressée au bureau d'inspection local de Consommation et Corporations Canada. Les exigences de marquage sont définies dans les articles allant de 18 à 26 du Règlement sur les poids et mesures. Les exigences relatives à l'installation, à l'utilisation et au mode d'emploi sont définies dans la partie V et dans les prescriptions établies en vertu de l'article 27 dudit règlement.




W.R. Virtue

Chief  
Legal Metrology Laboratories

Chef  
Laboratoires de la Métrologie légale

FILE/Dossier: O6922-P272-20  
PROJECT/Projet: AP-MI-86-0023

OCT 23 1986



	\$ 39	51	66	80	112	135	157	1.00	2.02	2.25	2.47
	\$ 40	105	165	3 50				0 95			
	\$ 39	56	65	112	141	169	197	2 25	2 53	2 81	3 10
	\$ 34	51	68	90	112	135	167	1 80	2 02	2 25	2 47

50%