



Measurement Canada  
An agency of Industry Canada

Mesures Canada  
Un organisme d'Industrie Canada

MAL No. - N° De LAM  
**MAL-S.WA-4081 Rev. 1**

Date: February 21, 2000

File number: 26922-AP-AM-99-0127

**MODIFICATION ACCEPTANCE LETTER (MAL)  
FOR  
Approved Weighing and Measuring Devices**

**Type of Device**

Electro-Mechanical Platform Scales

**Purpose**

The purpose of this letter is to convey the details of modifications to the range/classement for the weighing and measuring devices, which should have been previously listed for approval of type recognition from Measurement Canada. These modifications have been evaluated by the Approval Services Laboratory in accordance with national requirements and shall be considered within the scope of Notice of Approval S.WA-4081 and any subsequent revisions to this Notice of Approval.

**Applicant**

Bizerba Canada Inc.  
6419 Northam Drive  
Mississauga, Ontario  
L4V 1J2

**Description of Modifications**

Addition of the following dual-range model(s) 350:

0 - 15 kg x 0.005 kg

15 - 30 kg x 0.01 kg

0 - 30 kg x 0.01 kg

30 - 60 kg x 0.02 kg

0 - 60 kg x 0.02 kg

60 - 150 kg x 0.05 kg

Michel Maranda  
Approvals Technical Coordinator  
Mass Measurement Discipline



Mesures Canada  
Un organisme d'Industry Canada

Measurement Canada  
An agency of Industry Canada

MAL No. - N° De LAM  
**LAM-S.WA-4081 Rév. 1**

Date: 21 février 2000

Dossier: 26922-AP-AM-99-0127

**LETTRE D'ACCEPTATION DE MODIFICATION (LAM)**  
**pour**  
**Appareils de pesage et autres appareils de mesure approuvés**

**Type d'appareil**

Balances à plate-forme électromécaniques

**Objet**

La présente lettre vise à expliquer les détails des modifications apportées à l'étendue de mesure ou au classement des appareils de pesage et autres appareils de mesure, qui auraient dû être indiqués lors de l'approbation de type effectuée par Mesures Canada. Ces modifications ont été évaluées par le Laboratoire des services d'approbation en conformité aux exigences nationales et doivent être considérées comme faisant partie de la portée de l'avis d'approbation S.WA-4081 et de toutes ses révisions subséquentes.

**Requérant**

Bizerba Canada Inc.  
6419 Northam Drive  
Mississauga, Ontario  
L4V 1J2

**Descriptions des modifications**

Ajout des deux étendues de mesure suivantes pour le modèle 350 :

0 - 15 kg x 0.005 kg

15 - 30 kg x 0.01 kg

0 - 30 kg x 0.01 kg

30 - 60 kg x 0.02 kg

0 - 60 kg x 0.02 kg

60 - 150 kg x 0.05 kg

Michel Maranda  
Coordonnateur en technologie, approbations  
Discipline des masses