



Date: August 5, 2008

File number: AP-AV-08-0028

**MODIFICATION ACCEPTANCE LETTER (MAL)
FOR
Approved Measuring Devices**

Type of Device

Register

Purpose

The purpose of this letter is to convey details of modifications to liquid measuring devices which have previously received approval of type recognition from Measurement Canada. These modifications have been evaluated by the Approval Services Laboratory in accordance with national requirements and shall be considered to be included within the scope of the originally issued Notices of Approval specified in this document and any subsequent revisions to these Notices of Approval.

Scope

The modifications described in this letter affect devices identified in the following Notice(s) of Approval:

Approval Number

AV-2376

Applicant

VEEDER-ROOT
125 Powder Forest Drive
Post Office Box 2003
Simsbury, CT 06070-7684

Description of Modifications

To add the option of the use of approved and compatible mass flow meter, Actaris M200, with dual channel pulse output, used for volume measurement only.

Original copy signed by:

John Makin
Technical Coordinator – Liquid Measurement



Measurement Canada
An Agency of Industry Canada

Mesures Canada
Un organisme d'Industrie Canada

MAL No. - N° De LAM
LAM-V177

Date: 5 août, 2008

Dossier: AP-AV-08-0028

LETTRE D'ACCEPTATION DE MODIFICATION (LAM)
pour
Appareils de mesure approuvés

Type d'appareil

Enregistreur

Objet

La présente lettre vise à expliquer les détails des modifications apportées aux compteurs dont le type a déjà été approuvé par Mesures Canada. Ces modifications ont été évaluées par le Laboratoire des services d'approbation en conformité aux exigences nationales et doivent être considérées comme faisant partie de la portée des avis d'approbation d'origine énumérés dans le présent document et de toutes leurs révisions subséquentes.

Portée

Les modifications décrites dans la présente lettre s'appliquent aux compteurs visés par les avis d'approbation suivant :

Numéro d'approbation

AV-2376

Requérant

VEEDER-ROOT
125 Powder Forest Drive
Post Office Box 2003
Simsbury, CT 06070-7684

Description des modifications

Pour ajouter l'option de l'utilisation du débitmètre massique approuvé et compatible , Actaris M200, avec deux canaux de sortie, pour la mesure de volume seulement.

Copie authentique signée par :

John Makin
Coordonnateur technique – Mesure des liquides