



**MODIFICATION ACCEPTANCE LETTER  
(MAL)**

**for  
Approved Measuring Devices**

The purpose of this letter is to convey details of modifications to liquid measuring devices, which have previously received approval of type recognition from Measurement Canada. These modifications have been evaluated by the Approval Services Laboratory in accordance with national requirements and shall be considered to be included within the scope of the originally issued Notices of Approval specified in this document and any subsequent revisions to these Notices of Approval.

The modifications described in this letter affect devices identified in the following Notice(s) of Approval:

**LETTRE D'ACCEPTATION DE  
MODIFICATION (LAM)**

**pour  
Compteurs et appareils approuvés**

La présente lettre vise à expliquer les détails des modifications apportées aux appareils de mesure de liquides dont le type a déjà été approuvé par Mesures Canada. Ces modifications ont été évaluées par le Laboratoire des services d'approbation en conformité aux exigences nationales et doivent être considérées comme faisant partie de la portée des avis d'approbation d'origine énumérés dans le présent document et de toutes leurs révisions subséquentes.

Les modifications décrites dans la présente lettre s'appliquent aux appareils visés par les avis d'approbation suivants :

**APPROVAL NUMBER(S) | NUMÉRO(S) D'APPROBATION**

AV-2452

**TYPE OF DEVICE**

Electronic Register

**TYPE D'APPAREIL**

Enregistreur électronique

**APPLICANT**

Advanced Flow Solutions, Inc., d/b/a Liquid Controls  
9201 N I- 35 Service Road  
Oklahoma City, OK 73131, USA

**REQUÉRANT**

## DESCRIPTION OF MODIFICATIONS

The purpose of this modification is to add the following version 1.17.02 to the list of approved “LCR.iQ” software:

SR1000 v1.17.02,  
SR1010 v1.17.02,  
SR1011 v1.17.02,  
DT1000 v1.17.02.

The changes in this modification are non-metrological.

## DESCRIPTION DES MODIFICATIONS

Le but de cette modification est d'ajouter la version 1.17.02 ci-dessous à la liste des logiciels « LCR.iQ » approuvés :

SR1000 v1.17.02,  
SR1010 v1.17.02,  
SR1011 v1.17.02,  
DT1000 v1.17.02.

Les changements dans cette modification sont non métrologiques.

## SIGNATURE AND DATE

## SIGNATURE ET DATE

Original copy signed by: | Copie authentique signée par :

**Farhad Sharifi**  
(Acting) Manager, Liquids Measurement Laboratory  
(Intérim) Gestionnaire, Laboratoire de mesures des liquides

Modification Acceptance Letter issued on | Lettre d'acceptation de modification émise le :

**Date: 2024-05-13**

Web Site Address | Adresse du site Internet :  
<http://mc.ic.gc.ca>