



Measurement Canada
An agency of Industry Canada

Mesures Canada
Un organisme d'Industrie Canada

MAL No. - NE De LAM
MAL-V1 Rev. 2

Date: May 05 , 2001

File number: AP-AV-01-0004

**MODIFICATION ACCEPTANCE LETTER (MAL)
FOR
Approved Measuring Devices**

Type of Device

Fuel Dispenser

Purpose

The purpose of this letter is to convey details of modifications to liquid measuring devices which have previously received approval of type recognition from Measurement Canada. These modifications have been evaluated by the Approval Services Laboratory in accordance with national requirements and shall be considered to be included within the scope of the originally issued Notices of Approval specified in this document and any subsequent revisions to these Notices of Approval.

Revision 1

The purpose of Revision 2 is to correct the Approval Number from AV-2350 to AV- 2354. MAL-V1 Rev. 1 dated April 10, 2001. Please disregard MAL-V1 Rev. 1 dated April 10, 2001 that was previously issued. There is an error in the MAL.

Scope

The modifications described in this letter affect devices identified in the following Notice(s) of Approval:

Approval Number

AV-2202, AV-2203, AV-2206, AV-2207 and AV-2354

Applicant

Wayne Dresser
160 Cochrane Drive
Markham, Ontario
L3R 9S1

Description of Modifications

The dispensers approved under the above noted Notices of Approval can be equipped with electro-mechanical totalizers. The prefix "1" is added to the model number to indicate that the dispenser is equipped with these totalizers, (e.g. 1/V395D1).

Original copy signed by:

Randy Byrtus

Approvals Technical Coordinator
Fluid Measurement Discipline



Measurement Canada
An agency of Industry Canada

Mesures Canada
Un organisme d'Industrie Canada

MAL No. - NE De LAM
MAL-V1 Rév. 2

Date: 5 mai, 2001

Dossier: AP-AV-01-0004

LETTRE D'ACCEPTATION DE MODIFICATION (LAM) pour Compteurs et appareils approuvés

Type d'appareil

Distributeur de combustible

Objet

La présente lettre vise à expliquer les détails des modifications apportées aux compteurs dont le type a déjà été approuvé par Mesures Canada. Ces modifications ont été évaluées par le Laboratoire des services d'approbation en conformité aux exigences nationales et doivent être considérées comme faisant partie de la portée des avis d'approbation d'origine énumérés dans le présent document et de toutes leurs révisions subséquentes.

Révision 1

Le but de la révision 2 est de corriger le numéro d'approbation de AV-2350 à AV-2354. Si vous plaît, ne tenir aucun compte de MAL-V1 Rév. 1 daté 10 avril, 2001. Il ya une erreur dans le MAL.

Portée

Les modifications décrites dans la présente lettre s'appliquent aux compteurs visés par les avis d'approbation suivants:

Numéros d'approbation

AV-2202, AV-2203, AV-2206, AV-2207 et AV-2354

Requérant

Wayne Dresser
160 Cochrane Drive
Markham, Ontario
L3R 9S1

Description des modifications

Les distributeurs approuvés en vertu des avis d'approbation indiqués ci-dessus peuvent être équipés de totalisateurs électromécaniques. Le préfixe «1» est ajouté au numéro de modèle pour indiquer que le distributeur est muni de ces totalisateurs (ex. 1/V395D1).

Copie authentique signée par:

Randy Byrtus
Coordonnateur en technologie, approbations
Mesure des fluides