



Measurement Canada
An agency of Industry Canada

Mesures Canada
Un organisme d'Industrie Canada

MAL No. - N° De LAM
MAL-V9

Date: May 31, 1999

File number: AP-AV-99-0003

MODIFICATION ACCEPTANCE LETTER (MAL) FOR Approved Measuring Devices

Type of Device

Automatic Temperature Compensator

Purpose

The purpose of this letter is to convey details of modifications to liquid measuring devices which have previously received approval of type recognition from Measurement Canada. These modifications have been evaluated by the Approval Services Laboratory in accordance with national requirements and shall be considered to be included within the scope of the originally issued Notices of Approval specified in this document and any subsequent revisions to these Notices of Approval.

Scope

The modifications described in this letter affect devices identified in the following Notice(s) of Approval:

Approval Number

AV-2315 & AV-2346

Applicant

Wayne Dresser
160 Cochrane Drive
Markham, Ontario
L3R 9S1

Description of Modifications

The automatic temperature compensators, (ATCs), approved under the above noted Notices of Approval can be equipped with a thermistor type temperature sensor having part number 841-3060. This thermistor is the same as the previous thermistor used with these ATCs except that the connector on the end connecting to the potted harness has been changed from a round black circular connector to a white rectangular connector.

Randy Byrtus
Approvals Technical Coordinator
Fluid Measurement Discipline



Measurement Canada
An agency of Industry Canada

Mesures Canada
Un organisme d'Industrie Canada

MAL No. - N° De LAM

MAL-V9

Date: 31 mai 1999

Dossier: AP-AV-99-0003

LETTRE D'ACCEPTATION DE MODIFICATION (LAM)

pour Compteurs et appareils approuvés

Type d'appareil

Objet

La présente lettre vise à expliquer les détails des modifications apportées aux compteurs dont le type a déjà été approuvé par Mesures Canada. Ces modifications ont été évaluées par le Laboratoire des services d'approbation en conformité aux exigences nationales et doivent être considérées comme faisant partie de la portée des avis d'approbation d'origine énumérés dans le présent document et de toutes leurs révisions subséquentes.

Portée

Les modifications décrites dans la présente lettre s'appliquent aux compteurs visés par l'avis d'approbation suivant:

Numéro d'approbation

AV-2315 & AV-2346

Requérant

Wayne Dresser
160 Cochrane Drive
Markham, Ontario
L3R 9S1

Description des modifications

Les compensateurs automatiques de température (CAT) approuvés en vertu de l'avis d'approbation cité plus haut peuvent être équipés d'un capteur de température de type thermistance dont le numéro de pièce est 841-3060. Cette thermistance est la même que la thermistance précédemment utilisée avec ces CAT sauf que le connecteur à l'extrémité la reliant au câblage enrobé qui était un connecteur circulaire noir rond a été remplacé par un connecteur rectangulaire blanc.

Randy Byrtus
Coordonnateur en technologie, approbations
Mesure des fluides