



Measurement Canada
An agency of Industry Canada

Mesures
Un organisme d'Industrie Canada

MAL No. - N° De LAM
MAL-V86

Date: January 19, 2004

File number: AP-AV-03-0043

MODIFICATION ACCEPTANCE LETTER (MAL) FOR Approved Measuring Devices

Type of Device

Register

Purpose

The purpose of this letter is to convey details of modifications to liquid measuring devices which have previously received approval of type recognition from Measurement Canada. These modifications have been evaluated by the Approval Services Laboratory in accordance with national requirements and shall be considered to be included within the scope of the originally issued Notices of Approval specified in this document and any subsequent revisions to these Notices of Approval.

Scope

The modifications described in this letter affect devices identified in the following Notice(s) of Approval:

Approval Number

AV-2188 and AV-2342

Applicant

Liquid Controls, a Unit of IDEX Corporation
105 Albrecht Drive
Lake Bluff, Illinois, 60044-2242
USA

Description of Modifications

The devices approved under approvals AV-2188 and AV-2342 can use API Table 54C - Volume Correction Factors for Individual and Special Applications.

Original copy signed by:

Randy Byrtus
Technical Coordinator - Gas and Liquid Measurement



Measurement Canada
An agency of Industry Canada

Mesures Canada
Un organisme d'Industrie Canada

MAL No. - N° De LAM

MAL-V86

Date: 19 janvier 2004

Dossier: AP-AV-03-0043

LETTRE D'ACCEPTATION DE MODIFICATION (LAM)
pour
Compteurs et appareils approuvés

Type d'appareil

Enregistreur

Objet

La présente lettre vise à expliquer les détails des modifications apportées aux compteurs dont le type a déjà été approuvé par Mesures Canada. Ces modifications ont été évaluées par le Laboratoire des services d'approbation en conformité aux exigences nationales et doivent être considérées comme faisant partie de la portée des avis d'approbation d'origine énumérés dans le présent document et de toutes leurs révisions subséquentes.

Portée

Les modifications décrites dans la présente lettre s'appliquent aux compteurs visés par l'avis d'approbation suivant:

Numéro d'approbation

AV-2188 et AV-2342

Requérant

Liquid Controls, a Unit of IDEX Corporation
105 Albrecht Drive
Lake Bluff, Illinois, 60044-2242
USA

Description des modifications

Les appareils approuvés sous les approbations AV-2188 et AV-2342 peuvent employer la table 54C de l'API facteurs de volume pour les applications particulières et spéciales.

Copie authentique signée par :

Randy Byrtus

Coordonnateur technique - Mesure des gas et liquides