



Date: October 28, 2004

File number: AP-AV-04-0026

**MODIFICATION ACCEPTANCE LETTER (MAL)  
FOR  
Approved Measuring Devices**

**Type of Device**

Automatic Temperature Compensator

**Purpose**

The purpose of this letter is to convey details of modifications to liquid measuring devices which have previously received approval of type recognition from Measurement Canada. These modifications have been evaluated by the Approval Services Laboratory in accordance with national requirements and shall be considered to be included within the scope of the originally issued Notices of Approval specified in this document and any subsequent revisions to these Notices of Approval.

**Scope**

The modifications described in this letter affect devices identified in the following Notice(s) of Approval:

**Approval Number**

AV-2346

**Applicant**

Dresser Wayne, DI Canada Inc.  
40 Sharp road  
Brighton, Ontario  
K0K 1H0

**Description of Modifications**

To add the thermistor temperature probe model 888842-001 Rev B for use with the ATC modules approved under AV-2346.

Original signed by:

John Makin  
Technical Coordinator – Liquid Measurement



Date: 28 octobre 2004

Dossier: AP-AV-04-0026

## **LETTRE D'ACCEPTATION DE MODIFICATION (LAM)**

**pour**

**Appareils de mesure approuvés**

### **Type d'appareil**

Compensateur de température automatique

### **Objet**

La présente lettre vise à expliquer les détails des modifications apportées aux compteurs dont le type a déjà été approuvé par Mesures Canada. Ces modifications ont été évaluées par le Laboratoire des services d'approbation en conformité aux exigences nationales et doivent être considérées comme faisant partie de la portée des avis d'approbation d'origine énumérés dans le présent document et de toutes leurs révisions subséquentes.

### **Portée**

Les modifications décrites dans la présente lettre s'appliquent aux compteurs visés par l'avis d'approbation suivant :

### **Numéro d'approbation**

AV-2346

### **Requérant**

Dresser Wayne, DI Canada Inc.  
40 Sharp road  
Brighton, Ontario  
K0K 1H0

### **Description des modifications**

Pour ajouter le capteur de température de type thermistance du modèle 888842-001 Rev B pour utilisation avec les modules CTA approuvé par AV-2346.

Copie authentique signée par :

John Makin  
Coordonnateur technique – Mesure des liquides