



TRADE AND COMMERCE
CANADA

SD-A.251

STANDARDS DIVISION

OTTAWA,.....January 25, 1956.

APPROVAL LISTING

Bulk Meter Pump - Manufactured by Gilbert and Barker
Manufacturing Company, 829 Richmond Street West,
Toronto 1, Ontario.

Under the provisions of the Weights and Measures Act, Chapter 292,
R.S.C. 1952, and Regulations thereunder (P.C. 6894), the apparatus specified and
illustrated herein have been listed as approved devices and may be used in Canada
in accordance with the conditions applicable.

Apparatus Listed: Bulk meter pump, Models GPP201 and GPP202 with capacities of
13 G.P.M. and 18 G.P.M.

NOTE: These model numbers may be followed by the letter D and
numbers one to six. This suffix denotes electrical
characteristics of the motor.

Application: Metering wholesale quantities of gasoline or other petroleum products
for which it is calibrated.


Conditions: As prescribed in P.C. 6894, with the added mandatory requirement that
a metal plate be permanently affixed to the casing immediately below
the registering head, in view of the customer and bearing the words:-


NOT LEGAL FOR DELIVERY OF
QUANTITIES UNDER 40 GALLONS

The words: RESET COUNTER TO ZERO BEFORE EACH DELIVERY must appear
either on the counter or on the above-mentioned plate.

Description: These devices are meter pumps equipped with a non computing bulk
meter type registering head and may be equipped with a printer. The
mechanism is not equipped with an interlocking device as required on
retail meter pumps. The metering unit is similar to that used on
Gilbert and Barker meter pumps generally and a cylindrical sight glass
is installed outside the meter casing on the discharge line.

Testing: The standard tests for a meter pump shall apply as a preliminary test
and in addition, consecutive 5 gallon test runs shall be made to at
least a total of 40 gallons and the accumulated error compared with
the toleration of .25% in excess or deficiency.


C.S. Phillips,
Assistant Director (W&M),
Standards Division.


R.W. MacLean,
Director,
Standards Division.

Reference: A-366

SD-WA.251

GILBERT AND BARKER BULK METER PUMP

Model C17202

