



Measurement Canada
An Agency of Industry Canada

Mesures Canada
Un organisme d'Industrie Canada

MAL No. - N° De LAM
MAL-E271

Date: 2011-08-10

File number: 26681-AP-AE-11-0062

**MODIFICATION ACCEPTANCE LETTER (MAL)
FOR
Approved Meters and Devices**

Type of Device

Electricity Meters: Energy and Demand

Purpose

The purpose of this letter is to convey details of modifications to meters which have previously received approval of type recognition from An Agency of Industry Canada. These modifications have been evaluated by the Approval Services Laboratory in accordance with national requirements and shall be considered to be included within the scope of the originally issued Notices of Approval specified in this document and any subsequent revisions to these Notices of Approval.

Scope

The modifications described in this letter affect meters identified in the following Notices of Approval:

Approval Numbers


AE-1652

Applicant

General Electric
130 Main Street
Somersworth, New Hampshire, 03878 USA

Description of Modifications

Approval of I210+Ce meter model, which is I210+C with energy functions only.



Isabelle Tremblay
Laboratory Manager-
Electricity Measurement



Measurement Canada
An Agency of Industry Canada

Mesures Canada
Un organisme d'Industry Canada

MAL No. - N° De LAM
MAL-E271

Date: 2011-08-10

Dossier: 26681-AP-AE-11-0062

LETTRE D'ACCEPTATION DE MODIFICATION (LAM)

pour

Compteurs et Appareils Approuvés

Type d'Appareil

Compteurs d'électricité : Énergie et puissance

Objet

La présente lettre vise à expliquer les détails des modifications apportées aux compteurs dont le type a déjà été approuvé par Mesures Canada. Ces modifications ont été évaluées par le Laboratoire des services d'approbation en conformité aux exigences nationales et doivent être considérées comme faisant partie de la portée des avis d'approbation d'origine énumérés dans le présent document et de toutes leurs révisions subséquentes.

Portée

Les modifications décrites dans la présente lettre s'appliquent aux compteurs visés par les avis d'approbation suivant:

Numéros d'approbation


AE-1652

Requérant

General Electric
130 Main Street
Somersworth, New Hampshire, 03878 USA

Descriptions des modifications

Approbation du modèle de compteur I210
+ Ce, qui est le I210 + C avec fonctions
d'énergie seulement..



Isabelle Tremblay
Gestionnaire de laboratoire -
Mesure de l'électricité