



Consumer and
Corporate Affairs Canada

Consommation
et Corporations Canada

Legal Metrology
Branch

Direction de la
Métrologie légale

Holland Avenue
Ottawa, Ontario
K1A 0C9

Avenue Holland
Ottawa, Ontario
K1A 0C9

Your file Votre référence

Our file Notre référence

Telephone: (613)952-0640

File N°: 06565

Project:

Sangamo Canada
215 Laird Drive
Toronto, Ontario
M4G 3X1
Attn.: Mr. M. Purc

1988-12-15

Dear Mr. Purc,

This is in response to your recent notification given in accordance with the Electricity and Gas Inspection Regulations, Section 13, regarding devices approved under Notice of Approval E-60 to E-60-4, E-63 to E-63-3, E-64 to E-64-2, E-65 to E-65-8, E-66 to E-66-5, E-96 to E-96-3, E-137 to E-137-1. It is noted that these devices are to be modified as follows:

- New single-wall "P" base in place of original double-wall style;
- Stainless steel locking ring and boyonets (formerly aluminum);
- New resilient captive sealing gasket (formerly paper fibre);
- New sealing slider of hardened aluminum.

I am advised by the Electrical Metrology Laboratory that their evaluation of the proposed modifications indicates that they may be regarded as immaterial in relation to the approval provisions of the Act. Accordingly, your notification is acknowledged and the modifications are accepted as being included within the above-noted approval. All recipients of the approval notice are being informed by copy of this letter.

Yours truly,

W.R. Virtue, Chief
Legal Metrology Laboratories

cc: Approvals Mailing List