



Consumer and  
Corporate Affairs

Consommation et  
corporations

Standards

Normes

NOTICE OF APPROVAL  
AVIS D'APPROBATION

E - 105 - 1

Ottawa, May 18, 1976.

GENERAL ELECTRIC TYPE "D-51" IMPULSE GENERATOR

This Notice constitutes an amendment to Approval Notice E-105 dated July 9, 1971.

The output S.P.D.T. mercury wetted contact relay "Adams & Westlake AWCM-16182-1" has been replaced by the improved mercury wetted contact relay type "Midtex 160-261UA1, 9929515-5".

This modification improves the performance of type "D-51" impulse generator at lower temperature and permits the extension of the approved lower limit of ambient temperature from  $-23^{\circ}\text{C}$  to  $-30^{\circ}\text{C}$ .

Previous temperature range  $-23^{\circ}\text{C}$  to  $66^{\circ}\text{C}$  ( $-10^{\circ}\text{F}$  to  $150^{\circ}\text{F}$ ).

New temperature range  $-30^{\circ}\text{C}$  to  $66^{\circ}\text{C}$  ( $-22^{\circ}\text{F}$  to  $150^{\circ}\text{F}$ ).

Approval granted to:

Canadian General Electric  
Company Limited,  
1130 Boulevard Charest,  
Quebec, Quebec.

J. L. Armstrong, P.Eng.,  
Chief, Standards Laboratory,  
Metrology and Laboratory Services.

D. L. Smith, P.Eng.,  
Chief, Electricity and Gas Division,  
Metrology and Laboratory Services.

Ref: G 6565-C2-43