



Consumer and  
Corporate Affairs

Consommation et  
corporations

Standards

Normes

NOTICE OF APPROVAL  
AVIS D'APPROBATION

E-133

August 27, 1975  
Ottawa, \_\_\_\_\_

CANADIAN WESTINGHOUSE TYPE D4S AND D4A  
SINGLE PHASE TRANSFORMER TYPE WATTHOUR METERS

<u>Ampere Range</u>	<u>Volts</u>	<u>Kh</u>	<u>Wire</u>	<u>Register Ratio</u>
.12 - 10	120	0.3	2	333 1/3
.12 - 10	240	0.6	2	166 2/3
.12 - 10	240	0.6	3	166 2/3

Description

This line of transformer type meters is essentially the same, except for current conductors, as those approved under E-82 and E-82-1.

Transformer rated meters may also be offered, in which case custom register ratios will be used.

Approval granted to:

Westinghouse Canada Ltd.,  
Hamilton, Ontario.

J.L. Armstrong,  
Chief, Standards Laboratory,  
Metrology and Laboratory Services.

D.L. Smith,  
Chief, Electricity & Gas Division,  
Metrology and Laboratory Services.

Ref: GL 1145-57/C3-882