



Consumer and Corporate Affairs Canada Consommation et Corporations Canada

Legal Metrology Branch Direction de la Métrologie légale

Holland Avenue
Ottawa, Ontario
K1A 0C9

Avenue Holland
Ottawa, Ontario
K1A 0C9

Legal Metrology Laboratories
Telephone: (613) 952-0615

Your file Votre référence

Our file Notre référence

06565-E289
August 8, 1988
AML-E-42

Advanced Control Systems
P.O. Box 47824
Atlanta, Georgia
U.S.A. 30362
Attention: Mr. Jeff Chapman
EMAX Product Manager

Dear Mr. Chapman:


This is in response to your letter dated July 7, 1988, regarding Notices of Approval E-193 dated January 19, 1984 and issued to EMC Operations, Sparks, Maryland, and proposing to modify the Solid State Demand Recorder approved therein by substituting an ACS-designed modem board (PN 080015) on which the telephone interface is manufactured by Cermetek (Sunnyvale, California) and carries FCC Registration No. B468NR-68641-DP-E and the ringer equivalence of which is 0.2A, 0.5B.

I am advised that, upon evaluation of the proposed modifications, the Electrical Laboratory has concluded that they may be regarded as immaterial in relation to the approval provisions of the Electricity and Gas Inspection Act. Accordingly, pursuant to the authority delegated to me under the said Act:

1. your letter is hereby acknowledged as satisfying the notification requirements of section 13, Part III of the Electricity and Gas Inspection Regulations, SOR/86-131, 16 January, 1986, and
2. the proposed alteration is hereby accepted, in accordance with the aforementioned section, as being an immaterial one and hence included within the approval.

All recipients of Notices of Approval under the Electricity and Gas Inspection Act are being informed of this decision by copy of this letter.

Yours truly,


for W.R. Virtue
Chief

cc: Mailing List (Electrical Approvals)

Canada