



Industry and Science  
Canada

Industrie et Sciences  
Canada

Legal Metrology  
Branch

Direction de la  
Metrologie legale

Holland Avenue  
Ottawa, Ontario  
K1A 0C9

Avenue Holland  
Ottawa, Ontario  
K1A 0C9

Telephone: (613)952-0635

File N°: 06565-S771

Project: AP-EL-86-0007

AP-EL-86-0023

AP-EL-85-0033

AP-EL-87-0049

Scientific Columbus Company  
1900 Arlington Lane  
Columbus, Ohio  
USA 43228-4105  
Attention: Mr. Bill Jones

1993-09-23

Dear Mr. Jones:

This is in response to your recent notification given in accordance with the Electricity and Gas Inspection Regulations, Section 13, regarding devices approved under Notices of Approval E-199, E-214, E-218, E-219 and E-232. It is noted that these devices are approved as follows:

- Demand may be 20 minutes fixed (block) interval, or 20 minutes sliding window with a 1, 2, 4, 5 or 10 minute sub-interval.

I am advised by the Electrical Metrology Laboratory that its evaluation of your notification indicates that it is accepted as being included within the above-noted approvals. All recipients of the approval notices are being informed by copy of this letter.

Yours truly,

H.L. Fraser, Chief  
Electricity and Gas

cc: Approvals Mailing List