

Lanaua

Industrie Canada

Legal Metrology Branch Direction de la métrologie légale

Holland Avenue Ottawa, Ontario K1A OC9

Avenue Holland Ottawa (Ontario) K1A 0C9

Telephone: (613) 952-0635 FAX: (613) 952-1754 Téléphone: (613) 952-0635

File No.: 06565

Project: AP-EL-92-0022

Date: 1994-03-24

Schlumberger Industries Sangamo Electricity Division 215 Laird Drive Toronto, Ontario M4G 3X1 Attention: Mr. Marc Purc

Dear Mr. Purc:

This is in response to your recent notification given in accordance with the Electricity and Gas Inspection Regulations, Section 13, regarding devices approved under Notice of Approval E-227. It is noted that these devices may be modified as follows:

- 1. A new layout of the I/O board;
- 2. New terminal strips having provision for power input connections at both ends and rearranged input pulse connection sequence; and
- 3. A 26-conductor ribbon cable soldered to the I/O board.

I am advised by the Electrical Metrology Laboratory that its evaluation of the proposed modification indicates that it may be regarded as immaterial in relation to the approval provisions of the Act. Accordingly, your notification is acknowledged and the modification is accepted as being included within the above-noted approval. All recipients of the notice of approval are being informed by copy of this letter.

H.L. Fraser, Chief

Yours truly,

Electricity and Gas

cc: Approvals Mailing List

Canadä