



Consumer and
Corporate Affairs Canada

Consommation
et Corporations Canada

Legal Metrology
Branch

Direction de la
Métrologie légale

Holland Avenue
Ottawa, Ontario
K1A 0C9

Avenue Holland
Ottawa, Ontario
K1A 0C9

Your file *Voire référence*

Our file *Notre référence*

Telephone: (613)952-0640

File N°: 06565-S2
Project: AP-EL-88-0115

Sangamo Electricity Division
Schlumberger Canada Limited
215 Laird Drive
Toronto, Ontario
M4G 3X1

1989-04-26

Dear Mr. Purc,

This is in response to your recent notification given in accordance with the Electricity and Gas Inspection Regulations, Section 13, regarding devices approved under Notice of Approval E-240. It is noted that these devices are to be modified as follows:

- Magnetic reed switches changed to encapsulated configuration.
- Additional magnetic reed switch added to extended encoder circuit card so that switches Q1 and Q4 (which are angularly displaced about the 10-pole magnet) alternately actuate a set-reset latching circuit (U2) which has been added to the encoder card.

I am advised by the Electrical Metrology Laboratory that their evaluation of the proposed modifications indicates that they may be regarded as immaterial in relation to the approval provisions of the Act. Accordingly, your notification is acknowledged and the modifications are accepted as being included within the above-noted approval. All recipients of the approval notice are being informed by copy of this letter.

Yours truly,

W.R. Virtue, Chief
Legal Metrology Laboratories

cc: Approvals Mailing List

Canada