

DEPARTMENT OF TRADE AND COMMERCE  
Electrical and Gas Standards Laboratories


Ottawa, May 17th, 1932.

District Inspectors of Electricity & Gas;-

Dear Sirs,- Referring to National Research Council Circular #27 dated 30th ultimo regarding the G.9, GP.9 and GS.9 demand meters, manufactured by the Canadian General Electric Company, Toronto, Canada, I would advise that Form C of this type of meter has been approved by the National Research Council, and that the time interval of 10 minutes for the Form W and the Form C has also been approved by them, and may be admitted to verification in Canada.

The Form W refers to meters equipped with Warren motor drive and the form C refers to meters equipped with a clock drive.

Yours truly,



Director.