

DEPARTMENT OF TRADE AND COMMERCE  
ELECTRICITY AND GAS INSPECTION SERVICES

OTTAWA, March 2nd, 1934.

District Inspectors of Electricity & Gas:

Re - Landis & Gyr type CF4K, 2 & 3 wire,  
60 cycle prepayment watthour meter

Dear Sirs;-

I beg to advise you that the type CF4K  
2 and 3 wire 60 cycle prepayment watthour meter,  
manufactured by Landis & Gyr, Inc., Zug, Switzerland,  
and sold in Canada by Landis & Gyr, 104 Fifth Avenue,  
New York, has been approved by the National Research  
Council and may be admitted to verification in Canada  
in the following capacities;

Amperes - 5, 10, 15 and 25.  
Volts - 115 and 230.

I hope to be in a position to forward  
more detailed information regarding this meter in the  
near future.

Yours truly,

  
Director.

JLS/RS.

*60 cycle prepayment watthour meter*