



CANADA

DEPARTMENT OF TRADE AND COMMERCE
STANDARDS DIVISION

SD-EA.47

OTTAWA.....December 14, 1950.....

MODIFICATION OF TYPE APPROVAL

The modification to the apparatus specified herein has been duly approved by the Standards Division under the provisions of The Electricity Inspection Act, Chapter 22, 1928, as amended, and the modified apparatus may be admitted to verification in Canada.

Apparatus Modified: Type "L-30" Watthour Meters, manufactured by the Canadian General Electric Company, Limited, Toronto, Ontario.

Rating of Apparatus: All approved ratings of the above-named type of meter.

Modification: The modification, which is considered an immaterial change, consists of a new disc marking which includes the marks at present being applied. In addition to these present marks, the whole periphery of the disc is printed with two sets of marks, with 100 divisions on the outer circle and 130 divisions on the inner circle around the complete circumference. The marks on the inner circle may be used for stroboscopic testing, and the marks on the outer circle may be used to show the error of the meter directly in per cent, with some methods of calibrating. The present serrations on the edge of the disc will be retained for factory stroboscopic testing.

E. F. Power

E. F. Power,
Assistant Director, (E. & G.),
Standards Division.

R. W. Maclean

R. W. Maclean,
Director,
Standards Division.