

TRADE AND COMMERCE  
CANADA

## STANDARDS DIVISION

OTTAWA, November 10, 1954.

TYPE APPROVALFERRANTI TYPES "FMA" AND "FMS" SINGLE-PHASE WATTHOUR METERS

The apparatus specified herein has been duly approved by the Standards Division under the provisions of the Electricity Inspection Act, Chapter 94, R.S. 1952, and may be admitted to verification in Canada.

Apparatus Approved: Types "FMA" and "FMS" Single-Phase Watthour Meters, manufactured by Ferranti Electric Limited, Toronto, Ontario.

## Rating of Apparatus:

Nominal Amperes .....	5, 10, 15, 25, 50
Maximum Amperes .....	20, 40, 60, 100, 100
Volts .....	120, 240
Wire .....	2, 3, 2/3 (in 15 & 25 amperes)
Frequency .....	25, 60 cycles

Description: These meters are exactly the same as the types "FMA" and "FMS" approved by earlier circulars except that the nameplates are stamped with the 120- or 240-volt ratings. This approval permits the use of either the 115, 230 volt nameplates or the 120, 240 volt nameplates. Pertinent circulars are:-

SD-EA.22, dated May 23, 1950  
SD-EA.71, dated November 14, 1951  
SD-EA.111, dated July 17, 1953  
SD-EA.137, dated December 29, 1953.

E. F. Power,  
Assistant Director (E&G),  
Standards Division.

R. W. MacLean,  
Director,  
Standards Division.

Ref: A-406