



TRADE AND COMMERCE
CANADA

SD-EA.312

STANDARDS DIVISION

OTTAWA, September 17, 1957.

TYPE APPROVAL

SANGAMO TYPE "4L2" 3-PHASE, 4-WIRE Y DEMAND-ENERGY METER
FOR USE WITH INSTRUMENT TRANSFORMERS

The apparatus specified herein has been duly approved by the Standards Division under the provisions of the Electricity Inspection Act, Chapter 94, R.S. 1952, and may be admitted to verification in Canada.

Apparatus Approved: Type "4L2" 3-Phase, 4-Wire Y Demand-Energy Meter for use with instrument transformers, manufactured by the Sangamo Company Limited, Leaside, Toronto 17, Ontario.

Rating of Apparatus:

Current Range	0.12-8.5 amperes
Volts	115 or 120
Frequency	60 cycles
Demand Interval	20 and 32 minutes
Number of Elements	2 $\frac{1}{2}$
Kh	$\frac{1}{2}$
Demand Full Scale	3 KW
Type of Service	3-phase, 4-wire Y with instrument transformers.

Description: This approval extends that granted under SD-EA.266 of October 18, 1956, to this third rating for use with instrument transformers.

E. F. Power

E. F. Power,
Assistant Director (E&G),
Standards Division.

R. W. MacLean
R. W. MacLean,
Director,
Standards Division.

Ref: A-588A

