

AG-0310



Industry Canada
Legal Metrology Branch

Holland Avenue
Ottawa, Ontario
K1A 0C9

Telephone (613) 952-2638

Industrie Canada
Direction de la métrologie légale

Avenue Holland
Ottawa (Ontario)
K1A 0C9

Telephone (613) 952-2638

Project: AP-GL-95-0032

Date: 1996-01-15

EQUIMETER INC.
805 LIBERTY BOULEVARD
DUBOIS, PENNSYLVANIA
15801

Attn.: Mr. Paul Honchar

Dear Mr. Honchar

This is in response to your recent notification given in accordance with the Electricity and Gas Regulations, Section 13, regarding devices approved under Notice of Approval AG-0310. It is noted that these devices are to be modified as follows:

- 1) The material used to manufacture the index shaft worm gear the metric MR-12TC meters is changed from brass to a CELCON acetal copolymer plastic. The same plastic is used in the 415TC (imperial) meters. The metric index shaft worm gear blue in color in order to distinguish it from the white imperial 415TC gear.

I am advised by the Approval Technical Coordinator for the Gas Division that his evaluation of the proposed modifications indicates that they may be regarded as immaterial in relation to the approval provisions of the Act. Accordingly, your notification is acknowledged and the modifications are accepted as being included within the above-noted approval. All recipients of the approval notice are being informed by copy of this letter.

Yours truly,

Claude R. Bertrand, P.Eng
Manager, Approval Services Laboratory

cc: Approvals Mailing List

Canada



Industry Canada
Legal Metrology Branch

Holland Avenue
Ottawa, Ontario
K1A 0C9

Telephone (613) 952-2638

Industrie Canada
Direction de la métrologie légale

Avenue Holland
Ottawa (Ontario)
K1A 0C9

Telephone (613) 952-2638

Project: AP-GL-95-0032
Date: 1996-01-15

EQUIMETER INC.
805 LIBERTY BOULEVARD
DUBOIS, PENNSYLVANIA
15801
Att.: M. PAUL HONCHAR

Monsieur,

Nous accusons réception de l'avis de modification que vous nous avez envoyé récemment conformément à l'article 13 du Règlement sur l'inspection de l'électricité et du gaz, au sujet des appareils approuvés en vertu de l'avis d'approbation AG-0310. Il est à noter que ces appareils doivent être modifiés comme suit:

- 1) Le matériel de construction de l'engrenage à vis sans fin de l'arbre index des compteurs métriques MR-12TC a été changé du laiton à un plastique acetale copolymère CELCON. Le même plastique est utilisé pour les compteurs 415TC (impérial). Pour faciliter la différence entre les engrenages métriques et impériales, les engrenages métriques sont bleus.

Après avoir étudié les modifications proposées, le coordonnateur technique en approbations pour la discipline du gaz, me signale que ces modifications ne contreviennent en rien aux dispositions de la Loi. En conséquence, les modifications sont acceptées et intégrées de facto dans l'avis d'approbation susmentionné. Tous ceux qui ont reçu l'avis d'approbation sont avisés de cette décision par une copie de la présente lettre.

Je vous prie d'agréer mes salutations distinguées.

Claude Bertrand, ing.
Gérant, laboratoire des services d'approbation

cc: Liste d'envoi des approbations

Canada