



Industry and Science
Canada

Industrie et Sciences
Canada

Legal Metrology
Branch

Direction de la
métrologie légale

Holland Avenue
Ottawa, Ontario
K1A 0C9

Avenue Holland
Ottawa (Ontario)
K1A 0C9

Telephone: (613) 952-0835

Téléphone: (613) 952-0835

Project: AP-GL-93-0031
(Supersedes AP-GL-92-0055 dated 93/03/15)
Date: 1993-10-25

Mr. J.A. Waring
Dresser Canada Inc.
6688 Kitimat Road
Mississauga, Ontario
L5N 1P8

Dear Mr. Waring,

This is in response to your recent notification given in accordance with the Electricity and Gas Inspection Regulations, Section 13, regarding devices approved under Notices of Approval G-26 through G-26-14, and AG-0324. It is noted that these devices are to be modified as follows:

- TC register modules may now be equipped with a sealable adaptor plate and cross drilled screws, enabling the module to be verified in the meter shop, sealed, and installed, in-situ, on compatible, verified, and sealed Dresser meter modules.
- NOTE: The sealable register module, intended for interchangeability purposes, has a nameplate affixed containing the following information:
 - Manufacturer's name: Dresser Industries;
 - Model: TC
 - Serial number;
 - Part number;
 - Calibrated temperature range;
 - Departmental approval number; and
 - Conversion temperature.

I am advised by the Gas Metrology Laboratory that their evaluation of the proposed modifications indicates that they may be regarded as metrologically immaterial. Accordingly, your notification is acknowledged and the modifications are accepted as being included within the above-noted approval. All recipients of the approval notice are being informed by copy of this letter.

Yours truly,

H.L. Fraser, Chief
Electricity and Gas

Canada

cc: Approvals Mailing List