



Industry Canada
Legal Metrology Branch

Industrie Canada
Direction de la métrologie légale

Holland Avenue
Ottawa, Ontario
K1A 0C9

Avenue Holland
Ottawa (Ontario)
K1A 0C9

Telephone (613) 952-2638

Téléphone (613) 952-2638

Project: AP-GL-95-0031

Date: 1996-01-10

DRESSER CANADA INC.
6688 KITIMAT ROAD
MISSISSAUGA, ONTARIO
L5N 1P8

Attn.: Mr. Tony Waring

Dear Mr. Waring,

This is in response to your recent notification given in accordance with the Electricity and Gas Inspection Regulations, Section 13, regarding devices approved under Notice of Approval G-26, G-57, G-57 Rev.1 and 2, and AG-0324, AG-0324 Rev. 1 and 2. It is noted that these devices are to be modified as follows:

- 1) The material used to manufacture the gears used in the meters model LM-MA and L2-MA has been changed from brass to a DUPONT ZYTEL 101L lubricated polyanamide-nylon material. This gear assembly redirects the rotation of the accessory unit drive at a 90 degree angle. This change does not apply to the temperature compensated Instrument Drive or the High Pressure Planetary Instrument Drive units.

I am advised by the Approvals Technical Coordinator for the Gas Division that his evaluation of the proposed modifications indicates that they may be regarded as immaterial in relation to the approval provisions of the Act. Accordingly, your notification is acknowledged and the modifications are accepted as being included within the above-noted approvals. All recipients of the approval notices are being informed by copy of this letter.

Yours truly,

Claude R. Bertrand, P.Eng.
Manager, Approval Services Laboratory

cc: Approvals Mailing List

Canada



Industry Canada
Legal Metrology Branch

Industrie Canada
Direction de la métrologie légale

Holland Avenue
Ottawa, Ontario
K1A 0C9

Avenue Holland
Ottawa (Ontario)
K1A 0C9

Telephone: (613) 952-2638

Téléphone: (613) 952-2638

Project: AP-GL-95-0031

Date: 1996-01-10

DRESSER CANADA INC.
6688 KITIMAT ROAD
MISSISSAUGA, ONTARIO
L5N 1P8
Att.: M. Tony Waring

Monsieur,

Nous accusons réception de l'avis de modification que vous nous avez envoyé récemment conformément à l'article 13 du Règlement sur l'inspection de l'électricité et du gaz, au sujet des appareils approuvés en vertu des avis d'approbation G-26, G-57, G-57 rév.1 et 2, et AG-0324, AG-0324 rév. 1 et 2. Il est à noter que ces appareils doivent être modifiés comme suit:

- 1) Le matériel de construction des engrenages utilisées dans les compteurs modèles LM-MA et L2-MA a été changé du laiton à un matériel polyanamide-nylon lubrifié DUPONT ZYTEL 101L. Cet assemblage d'engrenages redirige à 90 degrés la rotation de l'unité motrice des accessoires. Ce changement n'affecte pas les compteurs à correction de température "Instrument Drive" ou les "Instrument Drive" planétaire à haute pression.

Après avoir étudié les modifications proposées, le coordonnateur technique en approbations pour la discipline du gaz, me signale que ces modifications ne contreviennent en rien aux dispositions de la Loi. En conséquence, les modifications sont acceptées et intégrées de facto dans l'avis d'approbation susmentionné. Tous ceux qui ont reçu l'avis d'approbation sont avisés de cette décision par une copie de la présente lettre.

Je vous prie d'agréer mes salutations distinguées.

Claude Bertrand, ing.
Gérant, laboratoire des services d'approbation

cc: Liste d'envoi des approbations

Canada