



*Legal Metrology
Branch
Holland Avenue
Ottawa, Ontario
K1A 0C9*

*Direction de la
Métrologie légale
Avenue Holland
Ottawa, Ontario
K1A 0C9*

Telephone: (613)952-0635

File No.: 06635
Project: AP-GL-93-0033
Date: 1993-10-28

JTT Barton Instruments Ltd.
3840 11A Street N.E.
Calgary, Alberta
T2E 6M6
Attn: Mr. Thomas Yoon

Dear Sirs:

This is in response to your recent notification given in accordance with the Electricity and Gas Inspection Regulations, Section 13, regarding devices approved under Notice of Approval AG-0339. It is noted that these devices are to be modified as follows:

- Addition of the following firmware versions:

N Gas B2.0.0S	N Gas B2.0.0C	N Gas S2.0.0S	N Gas S2.0.0C
N Gas M2.0.0S	N Gas M2.0.0C	N Gas B2.1.0S	N Gas B2.1.0C
N Gas S2.1.0S	N Gas S2.1.0C	N Gas M2.1.0S	N Gas M2.1.0C

The above firmware versions differ from those approved in AG-0339 in that they offer less flexibility in report generation and reduced access to communication networks.

I am advised by the Gas Metrology Laboratory that their evaluation of the proposed modifications indicates that they may be regarded as immaterial in relation to the approved provisions of the Act. Accordingly, your notification is acknowledged and the modifications are accepted as being included within the above-noted approval. All recipients of the approval notice are being informed by copy of this letter.

Yours truly,

H.L. Fraser, Chief
Electricity and Gas

cc. Approvals Mailing List



Industry Canada
Legal Metrology Branch

Industrie Canada
Direction de la métrologie légale

Holland Avenue
Ottawa, Ontario
K1A 0G9

Avenue Holland
Ottawa (Ontario)
K1A 0G9

Telephone: (613) 952-0835

Téléphone: (613) 952-0835

Projet: AP-GL-93-0033
Date: 1993-10-28

ITT Barton Instruments Ltd.
3840 - 11A Street, NE
Calgary, Alberta
T2E 6M6
Att.: Thomas Yoon

Monsieur,

Nous accusons réception de l'avis de modification que vous nous avez envoyé récemment conformément à l'article 13, du Règlement sur l'inspection de l'électricité et du gaz, au sujet des appareils approuvés en vertu de l'avis d'approbation AG-0339. Il est à noter que ces appareils doivent être modifiés comme suit:

- Les versions de microprogrammes suivantes sont ajoutées:
- | | | | |
|---------------|---------------|---------------|---------------|
| N Gas B2.0.0S | N Gas B2.0.0C | N Gas S2.0.0S | N Gas S2.0.0C |
| N Gas M2.0.0S | N Gas M2.0.0C | N Gas B2.1.0S | N Gas B2.1.0C |
| N Gas S2.1.0S | N Gas S2.1.0C | N Gas M2.1.0S | N Gas M2.1.0C |

Les versions de microprogrammation susmentionnées diffèrent de celles approuvées dans l'avis AG-0339 en présentant moins de souplesse dans la production de rapports et un accès réduit aux réseaux de communication.

Après avoir étudié les modifications proposées, le Laboratoire du gaz (métrologie) me signale que ces modifications ne contreviennent en rien aux dispositions de la Loi. En conséquence, les modifications sont acceptées et intégrées de facto dans l'avis d'approbation susmentionné. Tous ceux qui ont reçu l'avis d'approbation sont avisés de cette décision par une copie de la présente lettre.

Je vous prie d'agréer mes salutations distinguées.

H.L. Fraser, Chef
Électricité et gaz

cc: Liste d'envoi des approbations

Canada