



Measurement Canada  
An agency of Industry Canada

Mesures Canada  
Un organisme d'Industrie Canada

MAL No. - NE De LAM  
**MAL-G92**

Date: November 3, 2000

File number: AP-AG-00-0013

## **MODIFICATION ACCEPTANCE LETTER (MAL) FOR Approved Meters and Devices**

### **Type of Device**

Electronic Volume Transponder  
(Encoder/Receiver/ Transmitter)

### **Purpose**

The purpose of this letter is to convey details of modifications to meters which have previously received approval of type recognition from Measurement Canada. These modifications have been evaluated by the Approval Services Laboratory in accordance with national requirements and shall be considered to be included within the scope of the originally issued Notices of Approval specified in this document and any subsequent revisions to these Notices of Approval.

### **Scope**

The modifications described in this letter affect meters identified in the following Notice of Approval:

AG-0371

### **Applicant**

Itron Inc.  
2818 North Sullivan Road,  
Spokane WA 99216

### **Description of Modifications**

The housing of the Itron/Metscan model 40G Remote has increased in volume to about double the original size. It is now roughly 11.4 cm x 11.4cm x 2.5 cm. The colour has been changed from grey to black as well.

Original copy signed by:

Randy Byrtus  
Approvals Technical Coordinator  
Fluid Measurement Discipline



Measurement Canada      Mesures Canada  
An agency of Industry Canada    Un organisme d'Industrie Canada

MAL No. - NE De LAM  
**MAL-G92**

Date :                    3 novembre 2000

Dossier :                AP-AG-00-0013

**LETTRE D'ACCEPTATION DE MODIFICATION (LAM)**  
**pour**  
**Compteurs et appareils approuvés**

**Type d'appareil**

Transponder électronique de volume  
(encoder/récepteur/transmetteur)

**Objet**

La présente lettre vise à expliquer les détails de la modification apportée aux appareils dont le type a déjà été approuvé par Mesures Canada. Cette modification a été évaluée par le Laboratoire des services d'approbation en conformité aux exigences nationales et doit être considérée comme faisant partie de la portée des avis d'approbation d'origine et de toutes leurs révisions subséquentes.

**Portée**

Les modifications décrites dans la présente lettre s'appliquent aux compteurs visés par l'avis d'approbation suivant:

AG-0371

**Requérant**


Itron Inc.  
2818 North Sullivan Road,  
Spokane WA 99216

**Description des modifications**

Le taille du boîtier du modèle à distance 40G de Itron/Metscan a presque doublé par rapport à sa taille d'origine et mesure maintenant environ 11,4 cm sur 11,4 cm sur 2,5 cm. La couleur a également changé, du gris au noir.

Copie authentique signée par:

Randy Byrtus  
Coordonnateur technique d'approbations  
Discipline de la mesure des fluides



Itron 40G-Remote/Itron 40G-à distance