



Date: 2009-12-21

File number: AP-AG-09-0059

**MODIFICATION ACCEPTANCE LETTER (MAL)  
FOR  
Approved Meters and Devices**

**Type of Device**

Electronic Conversion Device

**Purpose**

The purpose of this letter is to convey details of modifications to meters which have previously received approval of type recognition from Measurement Canada. These modifications have been evaluated by the Approval Services Laboratory in accordance with national requirements and shall be considered to be included within the scope of the originally issued Notices of Approval specified in this document and any subsequent revisions to these Notices of Approval.

**Scope**

The modifications described in this letter affect meters identified in the following Notice (s) of Approval:

**Approval Number**

AG-0562

**Applicant**

Mercury Instruments  
3940 Virginia Ave.  
Cincinnati, Ohio  
45227

**Description of Modifications**

The purpose of this Modification Acceptance Letter is to add firmware version 1.07.00 to the Notice of Approval. This version increases the length of the alarm pulse to 6 minutes. No other changes were made.

**Original copy signed by:**

Randy Byrtus, CET  
Technical Coordinator - Gas Measurement



Date : 2009-12-21

Dossier : AP-AG-09-0059

**LETTRE D'ACCEPTATION DE MODIFICATION (LAM)**  
**pour**  
**Compteurs et appareils approuvés**

**Type d'appareil**

Correcteur électronique

**Objet**

La présente lettre vise à expliquer les détails des modifications apportées aux compteurs dont le type a déjà été approuvé par Mesures Canada. Ces modifications ont été évaluées par le Laboratoire des services d'approbation en conformité aux exigences nationales et doivent être considérées comme faisant partie de la portée de l'avis d'approbation d'origine énuméré dans le présent document et de toutes les révisions subséquentes.

**Portée**

Les modifications décrites dans la présente lettre s'appliquent aux compteurs visés par l'avis d'approbation suivant :

**Numéro d'approbation**

AG-0562

**Requérant**

Mercury Instruments  
3940 Virginia Ave.  
Cincinnati, Ohio  
45227

**Description des modifications**

Le but de cette lettre d'acceptation de modification est d'ajouter la version 1.07.00 du microprogramme à l'avis d'approbation. Cette version augmente la longueur de l'impulsion de l'alarme à 6 minutes. Aucun autre changement a été effectué.

**Copie authentique signée par:**  
Randy M. Byrtus, TEC  
Coordonnateur technique - Mesure des gaz