



Measurement Canada
An Agency of Industry Canada

Mesures Canada
Un organisme d'Industrie Canada

MAL No. - N° de LAM
MAL - G300

Date: 2011-07-18

File number: AP-AG-11-0024

MODIFICATION ACCEPTANCE LETTER (MAL) FOR Approved Meters and Devices

Type of Device

Electronic Pulse Recorder

Purpose

The purpose of this letter is to convey details of modifications to meters which have previously received approval of type recognition from Measurement Canada. These modifications have been evaluated by the Approval Services Laboratory in accordance with national requirements and shall be considered to be included within the scope of the originally issued Notices of Approval specified in this document and any subsequent revisions to these Notices of Approval.

Scope

The modifications described in this letter affect meters identified in the following Notice (s) of Approval:

Approval Numbers

AG-0594

Applicant

Mercury Instruments
3940 Virginia Ave.
Cincinnati, Ohio, 45227
USA

Description of Modifications

The electronic pulse recorder approved under AG-0594 can be equipped with firmware version 1.3. Also, the firmware version currently identified in approval AG-0594 should read 1.2 instead of 12.

Original copy signed by:

Randy Byrtus, CET
Manager, Gas Measurement



Measurement Canada
An Agency of Industry Canada

Mesures Canada
Un organisme d'Industrie Canada

MAL No. - N° De LAM
LAM - G300

Date : 2011-07-18

Dossier : AP-AG-11-0023

**LETTRE D'ACCEPTATION DE MODIFICATION (LAM)
pour
compteurs et appareils approuvés**

Type d'appareil

Enregistreur électronique d'impulsions

Objet

La présente lettre vise à expliquer en détail les modifications apportées aux compteurs dont le type a déjà été approuvé par Mesures Canada. Ces modifications ont été évaluées par le Laboratoire des services d'approbation conformément aux exigences nationales et doivent être considérées comme faisant partie de la portée des avis d'approbation d'origine énumérés dans le présent document et de toutes leurs révisions subséquentes.

Portée

Les modifications décrites dans la présente lettre s'appliquent aux compteurs visés par l'avis d'approbation suivant :

Numéro d'approbation

AG-0594

Requérant

Mercury Instruments
3940 Virginia Ave.
Cincinnati, Ohio, 45227
USA

Description des modifications

L'enregistreur électronique d'impulsions approuvé en vertu de l'avis AG-0594, peut être équipé de la version 1.3 du micrologiciel. De plus, la version du micrologiciel identifié présentement dans l'approbation AG-0594 devrait lire 1.2 au lieu de 1.3.

Copie authentique signée par :

Randy Byrtus, TEC
Gestionnaire, Mesure des gaz