



Measurement Canada
An Agency of Industry Canada

Mesures Canada
Un organisme d'Industrie Canada

MAL No. - N° De LAM
MAL - G193 Rev. 2

Date: March 1, 2006

File number: AP-AG-05-0042

MODIFICATION ACCEPTANCE LETTER (MAL) FOR Approved Meters and Devices

Type of Device

Gas Chromatograph

Purpose

The purpose of this letter is to convey details of modifications to meters which have previously received approval of type recognition from Measurement Canada. These modifications have been evaluated by the Approval Services Laboratory in accordance with national requirements and shall be considered to be included within the scope of the originally issued Notices of Approval specified in this document and any subsequent revisions to these Notices of Approval.

Scope

The modifications described in this letter affect meters identified in the following Notice (s) of Approval:

Approval Number

AG-0517

Applicant

Daniel Industries Canada Inc
A Division of Emerson Process Management
4215-72nd Avenue SE
Calgary, Alberta
T2C 2G5

Description of Modifications

The gas chromatographs, model 700, approved under the approval AG-0517 can be equipped with firmware version 2.04, 2.05 or 2.06. The model 700 gas chromatograph's firmware version can be viewed through the application / system menu.

Revision 1

The purpose of this revision was to upgrade the model 700 Gas Chromatograph firmware to include version 2.05.

Revision 2

The purpose of this revision is to upgrade the model 700 Gas Chromatograph firmware to include version 2.06.

Original signed by:

Randy Byrtus, CET
Approvals Technical Coordinator
Gas Measurement



Date: 1 mars 2006

Dossier: AP-AG-05-0042

LETTRE D'ACCEPTATION DE MODIFICATION (LAM)
pour
Compteurs et appareils approuvés

Type d'appareil

Chromatographe en Phase Gazeuse

Objet

La présente lettre vise à expliquer les détails des modifications apportées aux compteurs dont le type a déjà été approuvé par Mesures Canada. Ces modifications ont été évaluées par le Laboratoire des services d'approbation en conformité aux exigences nationales et doivent être considérées comme faisant partie de la portée des avis d'approbation d'origine énumérés dans le présent document et de toutes leurs révisions subséquentes.

Portée

Les modifications décrites dans la présente lettre s'appliquent aux compteurs visés par les avis d'approbation suivants :

Numéro d'approbation

AG-0517

Requérant

Daniel Industries Canada Inc
A Division of Emerson Process Management
4215-72nd Avenue SE
Calgary, Alberta
T2C 2G5

Description des modifications

Les chromatographes en phase gazeuse, modèle 700, approuvés en vertu de l'avis d'approbation AG-0517 peuvent être équipés avec le micrologiciel version 2.04, 2.05 ou 2.06. Le menu application / système, du chromatographe en phase gazeuse modèle 700, permet de voir la version de son microprogramme.

Révision 1

Le but de cette révision était pour faire une mise à niveau du micrologiciel, du Chromatographe modèle 700, afin d'y inclure la version 2.05.

Révision 2

Le but de cette révision est pour faire une mise à niveau du micrologiciel, du Chromatographe modèle 700, afin d'y inclure la version 2.06.

Copie authentique signée par :

Randy Byrtus, TEC
Coordonnateur Technique, Approbations
Mesure des Gaz