

M.O.J. Smith
W

Re: G-10

DEPARTMENT OF CONSUMER AND CORPORATE AFFAIRS

STANDARDS BRANCH

FILES: SL-100-6
SE-85-47
S-17-1

OTTAWA, September 16, 1971

TO: DISTRICT INSPECTORS OF ELECTRICITY AND GAS
FROM: CHIEF, ELECTRICITY AND GAS DIVISION

RE: APPROVAL BARTON MODEL 202A TRIPLE PEN RECORDER

It has been brought to our attention that Approval Notice G-10 does not specifically cover approval of single and two pen versions. Of course if all three pens in a three pen recorder can prove satisfactory there should be no doubt that a lesser number of pens should be alright as well. Although no mention was made in G-10, two earlier circulars have a bearing viz S-GA.233, June 18, 1962 and S-GA.232, June 18, 1962. These should be consulted.

Rather than re-issue circular G-10 at this time since we understand that Model 202A is in the process of being superseded, we suggest that a notation be made on G-10 copies to the effect that the two earlier circulars have a bearing and that single and two pen recorders are approved if properly marked.

W.J.S. Fraser

W.J.S. Fraser