



Consumer and
Corporate Affairs Canada

Legal Metrology

Consommation
et Corporations Canada

Métrologie légale

APPROVAL No. - N° D'APPROBATION

G-33-12

APR 5 1984
AVR

NOTICE OF APPROVAL

AVIS D'APPROBATION

Rockwell Series "G" Mark II Gas Turbo
Meters - External Lubrication.

Compteurs à gaz à turbine Rockwell Mark
II de la série "G" - système externe de
lubrification.

This Approval Notice supersedes Notices
of Approval G-33-10, G-33-2, G-33-3,
G-33-4, G-33-5, G-33-6, G-33-7, and
G-33-8.

Le présent avis remplace les avis
d'approbation G-33-10, G-33-2, G-33-3,
G-33-4, G-33-5, G-33-6, G-33-7, et
G-33-8.

Company: Rockwell International of
Canada Limited
41 Massey Road
P.O. Box 606
Guelph, Ontario
N1H 6L3

Société: Rockwell International of
Canada Limited
41 Massey Road
P.O. Box 606
Guelph, Ontario
N1H 6L3

Manufacturer: Rockwell International
of Canada Limited
Guelph, Ontario

Fabricant: Rockwell International of
Canada Limited
Guelph, Ontario

Type of Device: Rockwell Series "G"
Mark II, Gas Turbo Meters with readout/
output in Metric and Imperial units.

Type d'appareil: Compteurs Turbo pour
gaz, Rockwell, série "G" Mark II, avec
indicateur en unités métriques et
impériales.

Apparatus:

Appareil:

Model Designation:
RCM-T-18/MARK II T-18/MARK II
RCM-T-30/MARK II T-30/MARK II
RCM-T-60/MARK II T-60/MARK II
RCM-T-140/MARK II T-140/MARK II

Désignation de modèle:
RCM-T-18/MARK II T-18/MARK II
RCM-T-30/MARK II T-30/MARK II
RCM-T-60/MARK II T-60/MARK II
RCM-T-140/MARK II T-140/MARK II

.../2

Description: The models/types of Rockwell Turbo meters approved under this document are essentially identical to those previously approved except for the inclusion of an external lubricating system and the permanently lubricated type bearings have been replaced with bearings having shielding on only one side.

Terms and Conditions:

Manner of Use: These types of devices are permitted to be used in trade in accordance with sound measurement practice, installation and operating instructions provided by the manufacturer, terms and conditions appearing in the applicable Notices of Approval and applicable safety codes.

Reference No.: G6635-R2-25

Description: Les modèles/types de compteurs Turbo, Rockwell, approuvé sous ce document sont essentiellement identiques à ceux, déjà approuvé, sauf pour l'addition d'un système externe de lubrification, et le coussinet lubrifié à vie a été remplacé par un coussinet ayant qu'un seul coté blindé.

Conditions:

Utilisation: Ces types d'appareils peuvent être utilisés à des fins commerciales pourvu qu'ils le soit selon une méthode de mesurage fiable, qu'ils soit installés conformément aux prescriptions du fabricant et qu'ils satisfassent aux conditions du présent avis et des codes de sécurité pertinents.

No. de référence: G6635-R2-25



W.R. Virtue

Chief
Legal Metrology Laboratories

Chef
Laboratoires de la Métrologie légale

APR 5 1984