



Consumer and
Corporate Affairs

Consumption of
corporations

Standards

Normes

NOTICE OF APPROVAL
AVIS D'APPROBATION

G-40-3

Ottawa, October 3, 1975

CANADIAN METER COMPANY, BASE TEMPERATURE INDEX
TYPE I

This approval is supplementary to Circulars G-40, dated January 14, 1969 and G-40-2, dated July 5, 1973.


Apparatus


Canadian Meter Company Base Temperature Index, as covered under Notices of Approval G-40 and G-40-2. Except for the modification described below, all details of the apparatus, as outlined in the aforementioned Circulars, still apply.

Modification

The rear uncorrected counter has been relocated and is now facing the front. The position of the relocated counter is to the right of the existing corrected counter.

The terms of identification "FRONT" and "REAR" counters are no longer used. The counters are now identified by the terms "CORRECTED" and "UNCORRECTED".


J.L. Armstrong, P. Eng.,
Chief, Standards Laboratory,


D.L. Smith, P. Eng.,
Chief, Electricity & Gas Division,
Metrology and Laboratory Services.

Ref: GL 1147-57/C6-187