



Consumer and
Corporate Affairs Canada

Consommation
et Corporations Canada

Legal Metrology

Métrologie légale

APPROVAL No. - N° D'APPROBATION

G-44-5

NOV - 2 1983

NOTICE OF APPROVAL

AVIS D'APPROBATION

Sprague Meter Company Direct Reading
Digital Indexes

This approval supplements Notice of
Approval G-44, G-44-1, G-44-2 and
G-44-3.

Company: Sprague Meter, Division of
Textron Canada Ltd.,
85 Burford Road
Stoney Creek, Ontario
L8E 3C6

Manufacturer: Sprague Meter, Division
of Textron
Bridgeport, Connecticut
USA

Type of Device: Direct reading digital
indexes, part number 048203 for use on
Standard and Temperature Compensated
Meters.

APPARATUS

Part Number*: 048203

Faceplate Designation: 675 and 1000

Register Capacity, ft³: 9 999 900

No. of Registering Drums: 5

Test Dial Rotation, 10ft³: C.W.

Test Dial Rotation, 2½ft³: C.C.W.

Applicable Types of Gas Meters: 675
and 1000.

Indicateur numérique à lecture directe
de la Sprague Meter Company

Le présent avis complète les avis
d'approbation G-44, G-44-1, G-44-2 et
G-44-3.

Société: Sprague Meter, Division of
Textron Canada Limited
85 Burford Road
Stoney Creek, Ontario
L8E 3C6

Fabricant: Sprague Meter, Division
of Textron
Bridgeport, Connecticut
USA

Type d'appareil: Enregistreur numérique
à lecture directe, référence 048203 à
utiliser sur les compteurs standards et
à compensation de température.

APPAREIL

Référence*: 048203

Désignation du cadran: 675 et 1000.


Capacité de l'enregistreur, pi³:
9 999 900

Nombre de tambours indicateurs: 5

Rotation du cadran de contrôle, 10pi³:
horaire

Rotation du cadran de contrôle, 2½pi³:
anti-horaire

Types de compteurs de gaz applicables:
675 et 1000.



Description: These registers are of conventional design and are in addition to those Sprague Meter Company registers previously approved. All gears, union bearings, dial arms and odometer drums are made of a black molded plastic material. The index face is white ceramic on a copper backing, with black markings.

The faceplate designation number is located on the upper left corner of the faceplate.

* The part number is to appear on the register's gear frame.

All other provisions and stipulations contained in the Notice of Approval listed above remain in force.

Reference No.: G6635-S6-23

Description: Les présents enregistreurs sont de conception classique et s'ajoutent aux modèles de la Sprague Meter Company déjà approuvés. Tous les engrenages, les paliers raccords, les aiguilles du cadran et les tambours d'odomètre sont constitués d'une matière plastique moulée de couleur noire.

La référence du cadran se trouve dans le coin supérieur gauche du cadran.

*La référence doit figurer sur le support des engrenages de l'indicateur.

Toutes les autres dispositions et exigences indiquées dans l'avis susmentionné demeurent en vigueur.

No. de référence: G6635-S6-23


W.R. Virtue

Chief
Legal Metrology Laboratories

Chef
Laboratoires de la Métrologie légale

NOV - 2 1983