



Consumer and
Corporate Affairs

Consommation et
corporations

Standards

Normes

NOTICE OF APPROVAL
AVIS D'APPROBATION

G-56-2

Ottawa, March 31, 1976

MERCURY INSTRUMENTS, INC., MODELS MERCOR III - PT
AND MERCOR III - PTS VOLUME CORRECTING INTEGRATORS

This Approval is supplementary to Approval G-56-1
dated October 16, 1973.

Description

Approval has been granted for the inclusion of
two additional SUPPRESSED PRESSURE RANGES, VIZ. 200
TO 1200 PSIG AND 500 TO 1500 PSIG, to be used in the
above instruments. This Approval is for new instru-
ments only.

The inclusion of these two ranges involves only
changes to the pressure element and drum.

For all further information pertaining to these
devices, reference should be made to Approval Notice
G-56-1.

Approval granted to:

Baker Instruments, Limited,
Thornhill, Ontario &
Foothills Industrial Products,
Calgary, Alta.

J. L. Armstrong, P. Eng.,
Chief, Standards Laboratory,
Metrology and

D. L. Smith, P. Eng.,
Chief, Electricity & Gas Division
Laboratory Services

Ref: G 6635-M7#44

