



Consumer and  
Corporate Affairs

Consommation et  
corporations

Standards

Normes

**NOTICE OF APPROVAL  
AVIS D'APPROBATION**

**G-74-2**

Ottawa, October 28, 1976

**CANADIAN METER COMPANY CLOCK-TYPE PLASTIC  
REGISTERS FOR DOMESTIC GAS METERS**

This approval is supplementary to Notice of Approval G-74, dated January 20, 1971.

Apparatus

Register  
Designation

Dial Capacities,  
cu. ft. per rev.

Compatible Types  
of Gas Meters

Part No. 04972G072

1,000, 10,000, 100,000  
and 1,000,000 test dials  
2 and ½

AL175, ALC175  
A225, AL225, AL425

Description

The manufacturer introduced a new version of the plastic register. It is basically the same as the one previously approved under designation part No. 04972G037, except that the number 'B424' replaces the number 'B175' and the inscription 'Class 175-250' is no longer on the faceplate.

In other respects provisions of the Notice of Approval G-74 apply.

*J.L. Armstrong*  
J.L. Armstrong, P. Eng.  
Chief, Standards Laboratory,

*D.L. Smith*  
D.L. Smith, P. Eng.,  
Chief, Electricity and Gas Division,  
Metrology and Laboratory Services

Ref: G 6635-C6-23

