



Consumer and
Corporate Affairs Canada

Consommation
et Corporations Canada

Legal Metrology
Branch

Direction de la
Métrologie légale

Holland Avenue
Ottawa, Ontario
K1A 0C9

Avenue Holland
Ottawa, Ontario
K1A 0C9

Your file Votre référence

Telephone: (613) 990-8614

Our file Notre référence

O6635-M7-40
May 10, 1985
AML-G-16

Mercury Instruments Inc.
3940 Virginia Avenue
Cincinnati, Ohio
USA 45227

Attn: Mr. R. Comerford

Re: Application to Modify Approved Mercury Dual Instrument Drive Bracket,
Models 100 and 101.

Dear Mr. Comerford:

This letter is in response to your letter dated March 22, 1985, regarding a modification to the type/models of meters described above, which are currently approved in Notices of Approval G-80. I am advised that the Gas Laboratory has evaluated the modification to the bearing support system and using a single horizontal shaft instead of two. The conclusion is that it may be regarded as insignificant, in relation to the approval provisions of the Gas Inspection Act.

Accordingly, pursuant to the authority delegated to me by the Director under section 7 of the Gas Inspection Regulations;

1. your letter is hereby acknowledged as satisfying the notification requirements of subsection 7(2) of the regulations; and
2. the proposed alteration is hereby accepted, in accordance with subsection 7(3) of the Regulations, as being an immaterial one and hence included within the approval.

All recipients of Notices of Approval under the Gas Inspection Act are being informed of this decision by copy of this letter. For our records you are requested to advise us of the starting serial number when the change is initiated in production.

Yours truly,

W.R. Virtue
Chief,
Legal Metrology Laboratories

WRV:kl

cc: Mailing List (Gas Approvals)

Canada

