



Consumer and
Corporate Affairs

Consumption of
corporations

Standards

Normes

NOTICE OF APPROVAL
AVIS D'APPROBATION

G-83-1

Ottawa, July 31, 1975

PARKINSON COWAN (CANADA), TYPE B.250
POSITIVE DISPLACEMENT GAS METER

This is an amendment to Circular G-83.

The last sentence in the Notice of Approval G-83 which reads: "The stuffing boxes for the flag rods are presently equipped with additional packing grommets (flat neoprene washers) which help to stabilize meter's performance" is hereby revoked and replaced by: "The stuffing boxes are sealed by felt wadding".

Meters fitted with additional split or solid grommets in stuffing boxes shall not be permitted in service after December 31, 1976.

J.L. Armstrong
J.L. Armstrong,
Chief, Standards Laboratory,
Metrology and Laboratory Services.

D.L. Smith (for)
D.L. Smith,
Chief, Electricity & Gas Division,
Metrology and Laboratory Services.

Ref: GL 1147-57/P4-799

