



Legal Metrology Laboratories
Standards Building
Tunney's Pasture, Holland Avenue
Ottawa, Ontario
K1A 0C9
Telephone: (613) 995-0642

Votre référence Your file

Notre référence Our file

G-6635-1
February 28, 1984

TransCanada Pipelines
P.O. Box 54
Commerce Court West
Toronto, Ontario
M5L 1C2
Attn: Mr. R.A. Anzenars
Area Engineer

Re: Application to Modify Approved Rockwell Models TP-4 and TP-9 Turbine Gas Meters

Dear Mr. Anzenars:

This letter is in response to your letter of February 17, 1984 regarding a modification to the type/models of meters described above, which are currently approved in Notices of Approval G-98-1, G-98-2 and G-98-3. I am advised that the Gas Laboratory has evaluated the modification, that is, replacing the existing meter body thredolet (NPS $\frac{1}{2}$ F.S.) with a larger (NPS $\frac{1}{2}$ F.S.) thredolet, by welding and has concluded that it may be regarded as insignificant, in relation to the approval provisions of the Gas Inspection Act.

Accordingly, pursuant to the authority delegated to me by the Director under section 7 of the Gas Inspection Regulations;

1. your letter is hereby acknowledged as satisfying the notification requirements of subsection 7(s) of the regulations; and
2. the proposed alteration is hereby accepted, in accordance with subsection 7(3) of the Regulations, as being an immaterial one and hence included within the approval.

All recipients of Notices of Approval under the Gas Inspection Act are being informed of this decision by copy of this letter. For our records you are requested to advise us of the serial numbers of all such meters to be modified. In addition, you are required to notify the appropriate District Manager, Electricity and Gas Inspections, of the serial number and location, within a 30 day period of placing all such modified meters in service.

Yours truly

W.R. Virtue
Chief,
Legal Metrology Laboratories

cc: Mailing List (Gas Approvals)

Canada

