



CANADA

DEPARTMENT OF TRADE AND COMMERCE
STANDARDS DIVISION

SD-GA.5

OTTAWA, June 8, 1950.
(Superseding Cir. Letter SD-GC.12, dated Mar. 16, 1950)

MODIFICATION OF TYPE APPROVAL

The modification to the apparatus specified herein has been duly approved by the Standards Division under the provisions of the Gas Inspection Act, Chapter 82, R.S. 1927, and may be admitted to verification in Canada.

Apparatus Modified: Type "2B" Tin Gas Meter, manufactured by the American Meter Company, Albany, N.Y., U.S.A., and sold in Canada by the Canadian Meter Company, Hamilton, Ontario.

Modification: This modification (which is an amendment to Circular N.R.C.15 (Gas), dated April 19, 1947) approves the optional use of the ordinary all-brass 5-lt. side tubes in place of the Kerotest flared tube connections.

The use of the "2B" tin gas meter is not limited to the metering of propane in the bottled gas industry in Canada; it may be used for measuring any type of gas.

E. F. Power
E. F. Power,
Assistant Director (E&G),
Standards Division.

R. W. MacLean
R. W. MacLean,
Director,
Standards Division.