



CANADA

DEPARTMENT OF TRADE AND COMMERCE  
STANDARDS DIVISION

SD-GA.6

OTTAWA, August 16, 1950.

NOTIFICATION OF TYPE APPROVAL

The apparatus specified and illustrated herein has been duly approved by the Standards Division under the provisions of the Gas Inspection Act, Chapter 82, R.S. 1927, as amended, and may be admitted to verification in Canada.

Apparatus Approved: Type "2B-50" Gas Meter, manufactured by the American Meter Company, Albany, N.Y., and distributed in Canada by the Canadian Meter Company, Hamilton, Ontario.

Application: Measurement in distribution service of manufactured, natural, and petroleum gases, or mixtures thereof.

Rating of Apparatus: 50 cubic feet per hour.

Limiting Conditions: Approval granted provided that the words "per rev" be printed above the dials on the register to indicate the proper interpretation of the reading.

Description: This meter is of the positive displacement type and has a die cast aluminum case. It is 11½" high, 9" wide, and 8½" from front to back. The top casting, housing the valve mechanism, is bolted to the bottom casting which contains the bellows. A removable plate screwed to the top casting permits access to the adjustment. Facilities for sealing by wire and lead plug are provided by holes drilled in a number of screws holding the register window, the top plate, and the cover in place.

Testing: The standard tests for a meter of this rating will apply.

*W. J. S. Fraser*  
for E. F. Power,  
Assistant Director (E. & G.),  
Standards Division.

*R. W. MacLean*  
R. W. MacLean,  
Director,  
Standards Division.