



STANDARDS DIVISION

OTTAWA, November 19, 1952.

TYPE APPROVALFOXBORO CHARTS

The charts specified herein have been duly approved by the Standards Division under the provisions of the Gas Inspection Act, Chapter 82, R.S. 1927, as amended, and may be admitted to verification in Canada.

Charts Approved: The charts named on the attached list, manufactured by The Foxboro Company Limited, Ville LaSalle, P.Q. and The Foxboro Company, Foxboro, Mass., and distributed in Canada by Peacock Brothers Limited, Montreal, P.Q.

Description: The letter "M" after the identification number indicates that the chart has been printed in Canada.

All these charts are for use on Flow Meters and are of the "uniform" style, which means that they can be used with any instrument calibrated to the same uniform scale. A chart which is not "uniform" must be used with an instrument which has been specifically calibrated to that chart.

All charts, the identification number of which begins with "8" are called 12-inch charts, while those beginning with "7" are called 10-inch charts.

The symbol "SD" may appear on all charts, indicating that the charts have received Standards Division approval. However, this is not a mandatory requirement at the present time.

E. F. Power

E. F. Power,
Assistant Director (E&G),
Standards Division.

R. W. Maclean,
Director,
Standards Division.