

TRADE AND COMMERCE
CANADA

SD-GA.26

STANDARDS DIVISION

OTTAWA, December 12, 1952.

TYPE APPROVALBAILLEY CLASS 2 STATIC PRESSURE ELEMENT

The apparatus specified herein has been duly approved by the Standards Division under the provisions of The Gas Inspection Act, Chapter 82, R.S. 1927, as amended, and may be admitted to verification in Canada.

Apparatus Approved: Class 2 Static Pressure Element, manufactured by the Bailey Meter Company Limited, Montreal, P.Q.

Application: For measurement of static pressure of liquids or gases in the range of high pressures.

Rating of Apparatus:

Minimum Full Chart Range 0-60 p.s.i.

Maximum Full Chart Range 0-3000 p.s.i.

Description: The Class 2 Pressure Element is the type used by the manufacturer for high pressure service. It consists of a helical Bourdon tube having ten full turns. It is constructed of brass or stainless steel to suit specific conditions.

E. F. Power

E. F. Power,
Assistant Director (E&G),
Standards Division.

R. W. MacLean
R. W. MacLean,
Director,
Standards Division.

Ref: A-284