




ROCKWELL EMCO TYPES "150" AND "175" GAS METERS



CRANK BRACKET





Industry Canada
Legal Metrology Branch

Industrie Canada
Direction de la métrologie légale

Holland Avenue
Ottawa, Ontario
K1A 0C9

Avenue Holland
Ottawa (Ontario)
K1A 0C9

Telephone: (613) 952-0635

Téléphone: (613) 952-0635

Project: AP-GL-94-0007

Date: 1994-07-18

Argus Industrial Supply
154 Park Lane Avenue
Winnipeg, Manitoba
R2R 0K2
Attn.: Mr. Ross Roy

Dear Mr. Roy,

This is in response to your recent notification given in accordance with the Electricity and Gas Inspection Regulations, Section 13, regarding devices approved under:

Notice of Approvals:

G-23	G-44	G-99	G-127	G-153	S-GA-262
G-29	G-47	G-101	G-128	G-173	S-GA-302
G-32	G-60	G-117	G-133	G-198	SD-GA-30
G-37	G-86	G-125	G-138	AG-0292	SD-GA-108
G-43	G-94	G-126	G-145	AG-0311	SD-GA-148
				S-GA-247	SD-GA-153

Index covers called "lexan box assemblies" model: A-101, A-102, A-103; R-103, R-104, R-105, and R109; S-104, S-105, S-106, S-107 and its associated "Snap Seal" manufactured by Carolina Mouldings Inc. may be used in the meters approved under the above listed Notices of Approval.

Use of the above mentioned devices is also subject to the following conditions:


- (1) To facilitate verification, U.S. patent number "4621230" or the initials "CMI" shall be stamped on each index cover and shall be visible when the index cover is installed;

- 2 -

- (2) In the case where snap seals are used exclusively in sealing a meter, at least one seal, marked with the seal year and the meter verifier's identification, shall be applied to the point of access to the meter's adjustments.

I am advised by the Gas Metrology Laboratory that their evaluation of the proposed modifications indicates that they do not adversely affect the performance of the meter. Accordingly, your notification is acknowledged and the modifications are accepted as being included within the above-noted approval. All recipients of the approval notice are being informed by copy of this letter.

Yours truly,

A handwritten signature in black ink, appearing to read 'H.L. Fraser', written in a cursive style.

for H.L. Fraser, Director
Electricity and Gas

cc: Approvals Mailing List



Industry Canada
Legal Metrology Branch

Industrie Canada
Direction de la métrologie légale

Holland Avenue
Ottawa, Ontario
K1A 0C9

Avenue Holland
Ottawa (Ontario)
K1A 0C9

Telephone: (613) 952-0635

Téléphone: (613) 952-0635

Projet: AP-GL-94-0007

Date: 1994-07-18

Argus Industrial Supply
154 Park Lane Avenue
Winnipeg, Manitoba
R2R 0K2
Att.: M. Ross Roy

Monsieur,

Nous accusons réception de l'avis de modification que vous nous avez envoyé récemment conformément à l'article 13 du Règlement sur l'inspection de l'électricité et du gaz, au sujet des appareils approuvés en vertu des avis d'approbation suivants:

Avis d'approbation:

G-23	G-44	G-99	G-127	G-153	S-GA-262
G-29	G-47	G-101	G-128	G-173	S-GA-302
G-32	G-60	G-117	G-133	G-198	SD-GA-30
G-37	G-86	G-125	G-138	AG-0292	SD-GA-108
G-43	G-94	G-126	G-145	AG-0311	SD-GA-148
				S-GA-247	SD-GA-153

Les couvercles pour indicateurs désignés "boîtes lexan", modèles A-101, A-102, A-103, R-103, R-104, R-105, et R109, S-104, S-105, S-106, S-107 et les scellés emboîtables connexes de Carolina Mouldings Inc. peuvent être utilisés sur les compteurs visés par les avis d'approbation susmentionnés.

L'utilisation de ces dispositifs est également assujettie aux conditions suivantes :

- (1) Pour faciliter la vérification, le numéro de brevet É-U "4621230" ou les initiales "CMI" doivent être estampés sur chaque couvercle d'indicateur et doivent être visibles lorsque le couvercle est installé;
- (2) dans les cas où les scellés emboîtables sont utilisés exclusivement pour sceller un compteur, au moins un scellé, marqué de l'année de fabrication et de l'identification du vérificateur du compteur, doit être appliqué au point d'accès aux organes de réglage du compteur.

- 2 -

Après avoir étudié les modifications proposées, le Laboratoire du gaz (métrologie) me signale que ces modifications ne nuisent pas au rendement du compteur. En conséquence, les modifications sont acceptées et intégrées de facto dans les avis d'approbation susmentionnés. Tous ceux qui ont reçu l'avis d'approbation sont avisés de cette décision par une copie de la présente lettre.

Je vous prie d'agréer mes salutations distinguées.



H.L. Fraser, Directeur
Électricité et gaz

cc: Liste d'envoi des approbations