



TRADE AND COMMERCE
CANADA

STANDARDS DIVISION

OTTAWA, September 16, 1954.

TYPE APPROVAL

CANADIAN METER COMPANY TYPE "5B-225" ALUMINUMCASE GAS METER

The apparatus specified and illustrated herein has been duly approved by the Standards Division under the provisions of the Gas Inspection Act, Chapter 129, R.S. 1952, and may be admitted to verification in Canada.

Apparatus Approved: Type "5B-225" Aluminumcase Positive Displacement Gas Meter, manufactured by the Canadian Meter Company Limited, Hamilton, Ontario.

Rating of Apparatus:

Badged Capacity 170 cu.ft./hr.(air)
Differential Pressure
 at badged capacity 0.5" water gauge
Capacity per revolution111 cubic foot
Maximum Working Pressure 5 p.s.i.

Description: Except for the case, this meter is similar in construction to the Canadian Meter Company's ironcase meters. The index cover may be either of aluminum alloy and glass or of solid moulded plastic. Sealing arrangements are provided similar to those on the ironcase meter.

E. F. Power

E. F. Power,
Assistant Director (E&G),
Standards Division.

R. W. MacLean

R. W. MacLean,
Director,
Standards Division.

Ref: A-393