



TRADE AND COMMERCE  
CANADA

STANDARDS DIVISION

SD-GA.60

OTTAWA, December 21, 1954.

MODIFICATION OF TYPE APPROVAL

CANADIAN METER COMPANY TYPE "5B-225" ALUMINUMCASE GAS METER

The modification to the apparatus specified herein has been duly approved by the Standards Division under the provisions of the Gas Inspection Act, Chapter 129, R.S. 1952, and the modified apparatus may be admitted to verification in Canada.

Apparatus Modified: Type "5B-225" Aluminumcase Positive Displacement Gas Meter, manufactured by the Canadian Meter Company Limited, Hamilton, Ontario.

Rating of Apparatus:

- Badged Capacity ..... 170 cu.ft./hr.(air)
- Differential Pressure at badged capacity ..... 0.5" water gauge
- Capacity per revolution ..... .111 cu.ft.
- Maximum Working Pressure ..... 5 p.s.i.
- Connections (male) .....  $\frac{1}{2}$ ", 1" and 1 $\frac{1}{4}$ "

Modification: The modification permits the use of either light-weight aluminum or grey cast-iron meter tops, and also permits the three sizes of connections as listed above under 'Rating of Apparatus'. The meter is identical to the "5B-225" approved on September 16, 1954, under SD-GA.54.

E. F. Power,  
Assistant Director (E&G),  
Standards Division.

  
R. W. MacLean,  
Director,  
Standards Division.

Ref: A-393