



TRADE AND COMMERCE
CANADA

SD-GA.72

STANDARDS DIVISION

OTTAWA, July 7, 1955.

MODIFICATION OF TYPE APPROVAL

SPRAGUE TYPE "1A" 'ZEPHYR'
ALUMINUMCASE OR IRONCASE GAS METER

The modification to the apparatus specified herein has been duly approved by the Standards Division under the provisions of the Gas Inspection Act, Chapter 129, R.S. 1952, and the modified apparatus may be admitted to verification in Canada.

Apparatus Modified: Type "1A" Aluminumcase or Ironcase Positive Displacement Gas Meter, manufactured by the Sprague Meter Company, Bridgeport, Connecticut, U.S.A.

Rating of Apparatus:

Badged Capacity	140 cu.ft./hr.(air)
Differential Pressure at badged capacity	0.5" water gauge
Capacity per revolution	0.133 cu.ft.
Connections (male)	1"

Modification: The modification consists of re-rating the meter for "air". The meter is similar in construction to the meter approved under Circular 39/GAS of January 5, 1926. It may be supplied in either aluminum or iron case.

E. F. Power

E. F. Power,
Assistant Director (E&G),
Standards Division.

R. W. MacLean
R. W. MacLean,
Director,
Standards Division.