



TRADE AND COMMERCE
CANADA

SD-GA.77

STANDARDS DIVISION

OTTAWA, September 8, 1955.

TYPE APPROVAL

SPRAGUE TYPE "240"
COMBINATION METER AND REGULATOR

The apparatus specified and illustrated herein has been duly approved by the Standards Division under the provisions of the Gas Inspection Act, Chapter 129, R.S. 1952, and may be admitted to verification in Canada.

Apparatus Approved: Type "240" Combination Meter and Regulator, manufactured by the Sprague Meter Company, Bridgeport, Connecticut, U.S.A.

Application: Measurement in distribution service of manufactured, natural, and petroleum gases, or mixtures thereof.

Rating of Apparatus:

Badged Capacity 175 cu.ft./hr.(air)
Differential Pressure at badged capacity 0.5" water gauge
Capacity per revolution 0.133 cu.ft.

Description: The meter section is the same unit approved on Circular SD-GA.69, May 13, 1955. The regulator is a diaphragm type. The manufacturer has placed it in the top of the meter to provide economy in installation and better service on increased loads.

Note: Approval covers ironcase, aluminumcase, and iron top on aluminum castings. The approval also covers the meters in the standard and in the internal relief valve types.

E. F. Power

E. F. Power,
Assistant Director (E&G),
Standards Division.

R. W. MacLean
R. W. MacLean,
Director,
Standards Division.

Ref: A-480