



**Industry and Science  
Canada**

**Industrie et Sciences  
Canada**

**Legal Metrology  
Branch**

**Direction de la  
métrologie légale**

**Holland Avenue  
Ottawa, Ontario  
K1A 0C8**

**Avenue Holland  
Ottawa (Ontario)  
K1A 0C8**

**Telephone: (613) 952-0636**

**Téléphone: (613) 952-0636**

**Project: AP-GL-93-0030  
Date: 1993-10-25**

**Mr. G.F. Wilson  
Canadian Meter Company  
3037 Derry Road, West  
Milton, Ontario  
L9T 2X6  
Attn.: Glen F. Wilson**

**Dear Mr. Wilson,**

**This is in response to your recent notification given in accordance with the Electricity and Gas Inspection Regulations, Section 13, regarding devices approved under Notices of Approval SD-GA-89, SD-GA-122, G-23, G-125 and G-198. It is noted that these devices are to be modified as follows:**

- **Affixing of new die-cast hand hole cover with a cast receptacle to accommodate an encasement type seal.**

**I am advised by the Gas Metrology Laboratory that their evaluation of the proposed modifications indicates that they do not adversely affect the performance of the meter. Accordingly, your notification is acknowledged and the modifications are accepted as being included within the above-noted approval. All recipients of the approval notice are being informed by copy of this letter.**

**Yours truly,**

**H.L. Fraser, Chief  
Electricity and Gas**

**cc: Approvals Mailing List**

**Canada**

# HANDHOLE COVER AL-225/ AL-425

