



TRADE AND COMMERCE  
CANADA

SD-GA.102

STANDARDS DIVISION

OTTAWA, August 8, 1956.

TYPE APPROVAL

PARKINSON & COWAN TYPE "218"  
POSITIVE DISPLACEMENT GAS METER

The apparatus specified and illustrated herein has been duly approved by the Standards Division under the provisions of the Gas Inspection Act, Chapter 129, R.S. 1952, and may be admitted to verification in Canada.

Apparatus Approved: Type "218" Positive Displacement Gas Meter, manufactured by Parkinson & Cowan, London, England, and distributed in Canada by the Duragas Equipment Limited, 300 Greenwood Avenue, Toronto 6, Ontario.

Rating of Apparatus:

Badged Capacity .....	175 cu.ft./hr.(air)
Differential Pressure at badged capacity ....	0.5" water gauge
Capacity per revolution .....	.059 cu.ft.
Connections (male) .....	1" (6"centers)

Application: Measurement in distribution service of manufactured, natural, and petroleum gases, or mixtures thereof.

Description: This meter is identical internally to type "215" approved on Circular SD-GA.67, April 6, 1955. Modifications are:- Steel casing - center joint on case mechanically locked by a pressure toggle jig; hand hole for adjustment.

*E. F. Power*

E. F. Power,  
Assistant Director (E&G),  
Standards Division.

*R. W. MacLean*  
R. W. MacLean,  
Director,  
Standards Division.

Ref: A-582